

Upgrade® Disposable Prophy Angles

INDICATIONS FOR USE: Upgrade® Disposable Prophy Angles are single-use, disposable prophy angles designed to be used as part of a professionally delivered dental hygiene procedure. Upgrade® Disposable Prophy Angles can be used with any prophylaxis paste and will fit any Doriot-style dental handpiece that complies with ISO 3964.

CONTRAINDICATIONS: There are no known contraindications.

WARNING: Designed for use on low-speed air or electric handpieces. Do not operate at speeds exceeding 3,000 RPM. Use the lowest speed possible to achieve desired results.

PRECAUTIONS: Designed to be a single-use, disposable device. Dispose of the device after each patient use. Do not attempt to sterilize or re-use.

CAUTION: U.S. Federal law restricts this device to sale by or on the order of a dentist. For dental use only.

ADVERSE REACTIONS: In the present technical state, no adverse reactions have been reported so far.

OPERATING INSTRUCTIONS:

- Remove from packaging.
- Insert angle shaft into handpiece nose cone, ensuring slot on angle's shaft side aligns with pin on side of handpiece.
- Collet Grip: Tighten the chuck of the handpiece to engage the angle's shaft. Friction Grip: Push angle into nose cone firmly and gently give a small pull back to relieve axial pressure.
- Operate for a few seconds to ensure proper engagement of the shaft with the handpiece.
- Dispense prophy paste into the angle's cup and proceed with patient treatment. Do not operate angle at speeds exceeding 3,000 RPM.
- Dispose of angle after patient use.

0030502DF, R2-080714

Manufactured for:

Sultan Healthcare
1301 Smile Way • York, PA 17404-0807 • USA
Toll Free: 800-637-8582 • Phone: 201-871-1232
Fax: 201-871-0321 • www.sultanhealthcare.com

Made in U.S.A.

Upgrade® Disposable Prophy Angles

INDICATIONS FOR USE: Upgrade® Disposable Prophy Angles are single-use, disposable prophy angles designed to be used as part of a professionally delivered dental hygiene procedure. Upgrade® Disposable Prophy Angles can be used with any prophylaxis paste and will fit any Doriot-style dental handpiece that complies with ISO 3964.

CONTRAINDICATIONS: There are no known contraindications.

WARNING: Designed for use on low-speed air or electric handpieces. Do not operate at speeds exceeding 3,000 RPM. Use the lowest speed possible to achieve desired results.

PRECAUTIONS: Designed to be a single-use, disposable device. Dispose of the device after each patient use. Do not attempt to sterilize or re-use.

CAUTION: U.S. Federal law restricts this device to sale by or on the order of a dentist. For dental use only.

ADVERSE REACTIONS: In the present technical state, no adverse reactions have been reported so far.

OPERATING INSTRUCTIONS:

- Remove from packaging.
- Insert angle shaft into handpiece nose cone, ensuring slot on angle's shaft side aligns with pin on side of handpiece.
- Collet Grip: Tighten the chuck of the handpiece to engage the angle's shaft. Friction Grip: Push angle into nose cone firmly and gently give a small pull back to relieve axial pressure.
- Operate for a few seconds to ensure proper engagement of the shaft with the handpiece.
- Dispense prophy paste into the angle's cup and proceed with patient treatment. Do not operate angle at speeds exceeding 3,000 RPM.
- Dispose of angle after patient use.

0030502DF, R2-080714

Manufactured for:

Sultan Healthcare
1301 Smile Way • York, PA 17404-0807 • USA
Toll Free: 800-637-8582 • Phone: 201-871-1232
Fax: 201-871-0321 • www.sultanhealthcare.com

Made in U.S.A.

Upgrade® Disposable Prophy Angles

INDICATIONS FOR USE: Upgrade® Disposable Prophy Angles are single-use, disposable prophy angles designed to be used as part of a professionally delivered dental hygiene procedure. Upgrade® Disposable Prophy Angles can be used with any prophylaxis paste and will fit any Doriot-style dental handpiece that complies with ISO 3964.

CONTRAINDICATIONS: There are no known contraindications.

WARNING: Designed for use on low-speed air or electric handpieces. Do not operate at speeds exceeding 3,000 RPM. Use the lowest speed possible to achieve desired results.

PRECAUTIONS: Designed to be a single-use, disposable device. Dispose of the device after each patient use. Do not attempt to sterilize or re-use.

CAUTION: U.S. Federal law restricts this device to sale by or on the order of a dentist. For dental use only.

ADVERSE REACTIONS: In the present technical state, no adverse reactions have been reported so far.

OPERATING INSTRUCTIONS:

- Remove from packaging.
- Insert angle shaft into handpiece nose cone, ensuring slot on angle's shaft side aligns with pin on side of handpiece.
- Collet Grip: Tighten the chuck of the handpiece to engage the angle's shaft. Friction Grip: Push angle into nose cone firmly and gently give a small pull back to relieve axial pressure.
- Operate for a few seconds to ensure proper engagement of the shaft with the handpiece.
- Dispense prophy paste into the angle's cup and proceed with patient treatment. Do not operate angle at speeds exceeding 3,000 RPM.
- Dispose of angle after patient use.

0030502DF, R2-080714

Manufactured for:

Sultan Healthcare
1301 Smile Way • York, PA 17404-0807 • USA
Toll Free: 800-637-8582 • Phone: 201-871-1232
Fax: 201-871-0321 • www.sultanhealthcare.com

Made in U.S.A.

Upgrade® Disposable Prophy Angles

INDICATIONS FOR USE: Upgrade® Disposable Prophy Angles are single-use, disposable prophy angles designed to be used as part of a professionally delivered dental hygiene procedure. Upgrade® Disposable Prophy Angles can be used with any prophylaxis paste and will fit any Doriot-style dental handpiece that complies with ISO 3964.

CONTRAINDICATIONS: There are no known contraindications.

WARNING: Designed for use on low-speed air or electric handpieces. Do not operate at speeds exceeding 3,000 RPM. Use the lowest speed possible to achieve desired results.

PRECAUTIONS: Designed to be a single-use, disposable device. Dispose of the device after each patient use. Do not attempt to sterilize or re-use.

CAUTION: U.S. Federal law restricts this device to sale by or on the order of a dentist. For dental use only.

ADVERSE REACTIONS: In the present technical state, no adverse reactions have been reported so far.

OPERATING INSTRUCTIONS:

- Remove from packaging.
- Insert angle shaft into handpiece nose cone, ensuring slot on angle's shaft side aligns with pin on side of handpiece.
- Collet Grip: Tighten the chuck of the handpiece to engage the angle's shaft. Friction Grip: Push angle into nose cone firmly and gently give a small pull back to relieve axial pressure.
- Operate for a few seconds to ensure proper engagement of the shaft with the handpiece.
- Dispense prophy paste into the angle's cup and proceed with patient treatment. Do not operate angle at speeds exceeding 3,000 RPM.
- Dispose of angle after patient use.

0030502DF, R2-080714

Manufactured for:

Sultan Healthcare
1301 Smile Way • York, PA 17404-0807 • USA
Toll Free: 800-637-8582 • Phone: 201-871-1232
Fax: 201-871-0321 • www.sultanhealthcare.com

Made in U.S.A.